



CONTROL	L	N	H
---------	---	---	---

IMPORTANT: The barcode is for use only on the CELL-DYN Ruby. Refer to the appropriate System Operator's Manual for proper use of CELL-DYN Calibrator and Control Products.


IMPORTANT: Mixing and Handling

1. Remove a vial of the control from the refrigerator and warm to room temperature (18° to 30° C) for 15 minutes before use.

2. To mix: (**Do NOT mix mechanically or vortex.**)

For a video demonstration, visit www.corelaboratory.abbott and navigate to the Customer Portal → Technical Library → Other Reference Documents → Hematology Aids.

a. Hold the vial vertically and roll each vial between the palms of the hands for 15-20 seconds.



b. Continue to mix by holding the vial by the ends between the thumb and finger, rapidly inverting the vial 20 times end-over-end using a very quick turning motion of the wrist.



c. Analyze immediately after mixing. Subsequent analyses during this test period may be performed by inverting the vial 5 times prior to instrument analysis.

d. Steps a-c must be repeated upon removing the sample from the refrigerator for the entire open-vial time period regardless of the method of analysis (open tube, cap piercing, auto sample or manual sample).

3. Refer to the appropriate CELL-DYN System Operator's Manual for information about analyzing control specimens.

4. FOR AUTOMATED SAMPLING OR MANUAL CLOSED SAMPLING (CS):

- Refer to the appropriate CELL-DYN Operator's Manual. Remove the vial from the sample handler immediately after sampling.

FOR OPEN-VIAL SAMPLING:

- Aspirate a sample from the vial.
- Carefully wipe the vial rim and cap with a lint-free tissue.
- Replace the cap, ensuring it is on tight.

After sampling, return vial to refrigerator for maximum open-vial stability. If run in the open mode, wipe the threads of both vial and cap before replacing cap and returning to refrigerator.



2020-11-20

8 Consecutive Day Open-Vial Stability

		CONTROL L		CONTROL N		CONTROL H	
		LOT L0251		LOT N0251		LOT H0251	
SYSTEM	PARAMETER	ASSAY VALUE	± MEAN RANGE *	ASSAY VALUE	± MEAN RANGE *	ASSAY VALUE	± MEAN RANGE *
CELL-DYN 3700 SYSTEM	WOC 10 ⁹ /L	4.1	0.4	7.1	0.7	16.2	2.5
	WIC 10 ⁹ /L	4.1	0.5	7.1	1.0	16.3	3.0
	WBC 10 ⁹ /L	4.1	0.4	7.1	0.7	16.2	2.5
	NEU 10 ⁹ /L	2.3	0.3	4.0	0.8	9.2	2.0
	NEU %	56.9	8.0	56.8	8.0	56.8	10.0
	LYM 10 ⁹ /L	1.2	0.3	2.1	0.8	4.7	2.0
	LYM %	29.0	9.0	28.9	9.0	28.9	10.0
	MONO 10 ⁹ /L	0.4	0.2	0.7	0.4	1.6	0.6
	MONO %	9.8	5.0	9.9	5.0	9.9	3.0
	EOS 10 ⁹ /L	0.1	0.1	0.2	0.2	0.4	0.2
	EOS %	3.0	3.0	2.5	2.0	2.6	1.0
	BASO 10 ⁹ /L	0.1	0.1	0.2	0.2	0.6	0.6
	BASO %	3.0	3.0	3.0	3.0	3.0	3.0
	RBC 10 ¹² /L	2.87	0.15	4.13	0.20	5.13	0.30
	HGB g/dL	7.2	0.4	11.5	0.6	15.6	0.8
	HCT %	22.5	1.5	35.6	2.5	47.8	3.5
	MCV fL	78.3	4.0	86.1	4.0	93.2	5.0
MCH pg	25.1	2.0	27.8	2.0	30.4	2.0	
MCHC g/dL	32.0	2.3	32.3	3.0	32.6	2.3	
RDW %	20.3	2.5	21.3	2.5	17.5	2.5	
PLT 10 ⁹ /L	82	20	221	30	527	60	
MPV fL	8.2	3.0	7.9	3.0	7.6	3.0	
CELL-DYN Ruby SYSTEM	WBC (WOC) 10 ⁹ /L	4.0	0.4	7.1	0.7	16.2	2.5
	WBC (NOC) 10 ⁹ /L	4.0	0.4	7.1	1.0	16.5	2.5
	NEU 10 ⁹ /L	2.3	0.3	4.0	0.8	9.3	2.0
	NEU %	56.8	8.0	57.0	8.0	57.4	10.0
	LYM 10 ⁹ /L	1.1	0.3	2.0	0.8	4.6	2.0
	LYM %	28.7	9.0	28.8	9.0	28.6	10.0
	MONO 10 ⁹ /L	0.4	0.2	0.7	0.4	1.5	0.6
	MONO %	9.5	5.0	9.3	5.0	9.1	3.0
	EOS 10 ⁹ /L	0.1	0.1	0.2	0.2	0.4	0.2
	EOS %	3.0	3.0	2.5	2.0	2.5	1.0
	BASO 10 ⁹ /L	0.1	0.1	0.2	0.2	0.6	0.6
	BASO %	3.0	3.0	3.0	3.0	3.0	3.0
	RBC 10 ¹² /L	2.83	0.15	4.17	0.20	5.24	0.30
	HGB g/dL	7.1	0.4	11.4	0.6	15.6	0.8
	HCT %	20.1	1.5	31.7	2.5	42.2	3.5
	MCV fL	71.1	4.0	76.1	4.0	80.5	5.0
	MCH pg	25.1	2.0	27.3	2.0	29.8	2.0
MCHC g/dL	35.3	2.3	36.0	3.0	37.0	2.3	
RDW %	14.1	2.5	14.9	2.5	11.6	2.5	
PLT 10 ⁹ /L	72	20	204	30	504	60	
MPV fL	10.5	3.0	10.2	3.0	10.2	3.0	

* The **MEAN RANGE** does not represent standard deviations (SD).
NOTE: Flags may occur with control materials and should be disregarded.



CELL-DYN and CELL-DYN Ruby are trademarks of Abbott Laboratories in various jurisdictions.



Abbott Laboratories
Diagnostics Division
Abbott Park, IL 60064
USA



Abbott GmbH & Co. KG
Max-Planck-Ring 2
65205 Wiesbaden
Germany
+49-6122-580

MANUFACTURED FOR
Abbott Laboratories



REF 08H59-01, 08H59-02

