



CONTROL L N H


IMPORTANT: The barcode is for use only on the CELL-DYN Ruby. Refer to the appropriate System Operator's Manual for proper use of CELL-DYN Calibrator and Control Products.

IMPORTANT: Mixing and Handling


1. Remove a vial of the control from the refrigerator and warm to room temperature (18° to 30° C) for 15 minutes before use.
2. To mix: **(Do NOT mix mechanically or vortex.)**

For a video demonstration, visit www.corelaboratory.abbott and navigate to the Customer Portal → Technical Library → Other Reference Documents → Hematology Aids.

- a. Hold the vial vertically and roll each vial between the palms of the hands for 15-20 seconds.



- b. Continue to mix by holding the vial by the ends between the thumb and finger, rapidly inverting the vial 20 times end-over-end using a very quick turning motion of the wrist.



- c. Analyze immediately after mixing. Subsequent analyses during this test period may be performed by inverting the vial 5 times prior to instrument analysis.
- d. Steps a-c must be repeated upon removing the sample from the refrigerator for the entire open-vial time period regardless of the method of analysis (open tube, cap piercing, auto sample or manual sample).

3. Refer to the appropriate CELL-DYN System Operator's Manual for information about analyzing control specimens.

4. FOR AUTOMATED SAMPLING OR MANUAL CLOSED SAMPLING (CS):

- Refer to the appropriate CELL-DYN Operator's Manual. Remove the vial from the sample handler immediately after sampling.

FOR OPEN-VIAL SAMPLING:

- Aspirate a sample from the vial.
- Carefully wipe the vial rim and cap with a lint-free tissue.
- Replace the cap, ensuring it is on tight.

After sampling, return vial to refrigerator for maximum open-vial stability. If run in the open mode, wipe the threads of both vial and cap before replacing cap and returning to refrigerator.

		Exp. 2026-05-01		8 Consecutive Day Open-Vial Stability			
		CONTROL L		CONTROL N		CONTROL H	
		LOT	L6047	LOT	N6047	LOT	H6047
SYSTEM	PARAMETER	ASSAY VALUE	± MEAN RANGE *	ASSAY VALUE	± MEAN RANGE *	ASSAY VALUE	± MEAN RANGE *
CELL-DYN Ruby SYSTEM	WBC (WOC) 10 ⁹ /L	4.3	0.4	7.2	0.7	16.5	2.5
	WBC (NOC) 10 ⁹ /L	4.3	0.4	7.3	1.0	16.7	2.5
	NEU 10 ⁹ /L	2.5	0.3	4.2	0.8	9.8	2.0
	NEU %	58.8	8.0	58.6	8.0	59.2	10.0
	LYM 10 ⁹ /L	1.1	0.3	1.9	0.8	4.3	2.0
	LYM %	26.0	9.0	26.4	9.0	26.0	10.0
	MONO 10 ⁹ /L	0.4	0.2	0.7	0.4	1.5	0.6
	MONO %	9.2	5.0	9.2	5.0	8.8	3.0
	EOS 10 ⁹ /L	0.1	0.1	0.2	0.2	0.4	0.2
	EOS %	3.0	3.0	2.3	2.0	2.3	1.0
	BASO 10 ⁹ /L	0.2	0.1	0.3	0.2	0.6	0.6
	BASO %	3.7	3.0	3.6	3.0	3.7	3.0
	RBC 10 ¹² /L	2.85	0.15	4.19	0.20	5.26	0.30
	HGB g/dL	6.6	0.4	11.5	0.6	15.8	0.8
	HCT %	19.1	1.5	31.4	2.5	42.2	3.5
	MCV fL	67.1	4.0	75.0	4.0	80.3	5.0
	MCH pg	23.2	2.0	27.4	2.0	30.0	2.0
	MCHC g/dL	34.6	2.3	36.6	3.0	37.4	2.3
	RDW %	14.7	2.5	12.6	2.5	11.9	2.5
	PLT 10 ⁹ /L	78	20	229	30	539	60
MPV fL	10.1	3.0	9.7	3.0	9.8	3.0	

* The **MEAN RANGE** does not represent standard deviations (SD).
NOTE: Flags may occur with control materials and should be disregarded.



CELL-DYN and CELL-DYN Ruby are trademarks of Abbott Laboratories in various jurisdictions.

Abbott Laboratories
Diagnostics Division
Abbott Park, IL 60064
USA

EC REP Abbott GmbH & Co. KG
Max-Planck-Ring 2
65205 Wiesbaden
Germany
+49-6122-580

MANUFACTURED FOR
Abbott Laboratories



REF 08H59-01, 08H59-02

